

# Certificate of Conformance

**Company:** INSERT COMPANY NAME  
**Certificate Number:** SC02.15.19Z  
**Certification Date:** February 15th, 2019  
**Expiration Date:** February 15th, 2020



1295 Morningside Avenue, Unit 16-17  
Toronto, ON M1B 4Z4 Canada  
Phone: 416-261-4865 Fax: 416-261-7879  
www.scigiene.com

This product was Calibrated and Certified during manufacturing to be traceable to National Institute of Standards and Technology standards as per the following validation information.

**Test equipment used by Scigiene for verifying specifications and accuracy during calibration procedures:**

**Anova Model: A40 Refrigerated & Heating Circulator Bath** Serial#: 96453

**Anova I-100 Standard Immersion Circulator** Serial#: 65498

**Techne Model TE-10D**, Serial#: 174035-15

**Reference Standard**

**Eutechnics Precision Temperature Thermometer**, Model#: 4500 with SP646, Serial#: 10402616, -40°C to +150°C, 0.01 Resolution, Accuracy  $\pm 0.05^\circ\text{C}$  Calibration System Date: November 21st, 2018 Due Date: November 21st, 2019

**Model Number:** PTT450-GC - Green Pocket Thin Tip Quick Read  
**Waterproof Thermometer -40°C to +230°C**  
*The units has been tested at 0°C  
and found to be within  $\pm 0.5^\circ\text{C}$*

**Serial Number:**

11111			
22222			
33333			

Technician: Harvi Ancheta

A handwritten signature in black ink that reads "H.A. Ancheta".