

# Certificate of Traceability and Calibration

Company: Insert Company Name  
Certificate Number: SC10.04.19D  
Certification Date: October 4th, 2019  
Expiration Date: October 4th, 2020



1295 Morningside Ave., Unit 16-18  
Scarborough, ON. M1B 4Z4 Canada  
Phone: 416-261-4865 Fax: 416-261-7879

This product was Calibrated and Certified to be traceable to National Institute of Standards and Technology standards as per the following validation information.

**Test equipment used by Scigiene for verifying specifications and accuracy during calibration procedures:**

Hygiena EnSure Luminometer, Serial Number: 105135 with:

**Calibration Control Kit:** CalCheck LED Calibration Verification

Lot Number: 143718

Expiration: October 10th, 2021

## Using EnSure Meter

Model #	Serial #	Device Reading RLU	Specification Acceptance
ATP4001 - SCIGIENE CERTIFIED CALIBRATION ROD KIT	SC093016A	99	80-160
		0	0-4

Technician: Harvi Ancheta

A handwritten signature in black ink that reads "Harvi Ancheta".