

	<p style="text-align: center;"><b>System Sure Plus</b></p> 	<p style="text-align: center;"><b>EnSure</b></p> 
 <p style="text-align: center;"><b>UltraSnap</b></p>	<p style="text-align: center;">1 RLU = 1 fmol ATP</p> <p style="text-align: center;"><b>Pass &lt;10; Caution 11-29; Fail 30+</b></p>	<p style="text-align: center;">1 RLU = 0.5 fmol ATP</p> <p style="text-align: center;"><b>Pass &lt;20; Caution 21-59; Fail 60+</b></p>
 <p style="text-align: center;"><b>SuperSnap</b></p>	<p style="text-align: center;">1 RLU = 0.2 fmol ATP</p> <p style="text-align: center;"><b>Pass &lt;40; Caution 41-79; Fail 80+</b></p>	<p style="text-align: center;">1 RLU = 0.1 fmol ATP</p> <p style="text-align: center;"><b>Pass &lt;80; Caution 81-159; Fail 160+</b></p>



1295 Morningside Avenue, Unit 16-18  
 Scarborough, ON M1B 4Z4 Canada  
 Phone: 416-261-4865 Fax: 416-261-7879  
[www.scigiene.com](http://www.scigiene.com)